MALLORY Mallory Sonalert Products,Inc. Sales Outline Drawing			Part #: SBT12MMCPC  Revision: D		
					Specifications:
Sound Level Category:	Silence	Loud	Loud	Loud	
Mode of Operation:	Silence - NO SOUND	Low Priority Alarm 2 beeps repeating every 30 sec	Medium Priority Alarm 3 beeps repeating every 7.5 sec	High Priority Alarm 10 beeps repeating every 2.5 sec	
Loudness @ 10cm:		85 to 95 dB(A) Typ.	85 to 95 dB(A) Typ.	85 to 95 dB(A) Typ.	
Frequency:		150 to 4000 Hz			
Connection:	Connect "+" to "+" & "-" to "-" of Power Supply		Connect "+" to "+" & "-" &	Connect "+" to "+" & "-" & H to "-" of Power Supply	
Voltage Rating:		9 to 12 VDC			
Average Current Draw:	<15mA Typ.	<15mA Typ.	<15mA Typ.	<20mA Typ.	
Ipk Current*:		75mA			
Housing Material:		ABS 777D Plastic Resin, Black			
Storage Temperature:		-40° to +85° C			
Operating Temperature		-30° to +65° C			
Weight Typical:		0.6 oz (17g)			
Options:		Please contact factory.			
		CUL Recognized IT NUMBER: 7880593		ROHS Compliant	
Before Code Date:	1546 Not cUL Recognized	After Co	ode Date: 1547	cUL Recognized	
	Ø1.750 [Ø	Ø44.5]	0.334 [8.5]		
	0.100 [:	0.200 [5.1		Mode in U.S.A. MANURY  MANURY  Medical	

0.551 [14.0] -

\*Ipk = Instantaneous Peak

Ø0.039 [Ø1.0] 5 Holes Req'd

0.229 [5.8]